

Award Number:  
W81XWH-10-1-0573

TITLE:  
A prosthesis to train the proprioceptive capabilities of the residual limb  
of military personnel recovering from lower limb amputation

PRINCIPAL INVESTIGATOR:  
Brian Glaister

CONTRACTING ORGANIZATION:  
Cadence Biomedical  
Seattle, WA 98103

REPORT DATE:  
September 2012

TYPE OF REPORT:  
Annual

PREPARED FOR: U.S. Army Medical Research and Materiel Command  
Fort Detrick, Maryland 21702-5012

DISTRIBUTION STATEMENT: (Check one)

- ☒ Approved for public release; distribution unlimited
- ☐ Distribution limited to U.S. Government agencies only;  
report contains proprietary information

The views, opinions and/or findings contained in this report are those of the  
author(s) and should not be construed as an official Department of the Army  
position, policy or decision unless so designated by other documentation.

# REPORT DOCUMENTATION PAGE

Form Approved  
OMB No. 0704-0188

Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing this collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to Department of Defense, Washington Headquarters Services, Directorate for Information Operations and Reports (0704-0188), 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302. Respondents should be aware that notwithstanding any other provision of law, no person shall be subject to any penalty for failing to comply with a collection of information if it does not display a currently valid OMB control number. **PLEASE DO NOT RETURN YOUR FORM TO THE ABOVE ADDRESS.**

<b>1. REPORT DATE (DD-MM-YYYY)</b> September 2012		<b>2. REPORT TYPE</b> Annual		<b>3. DATES COVERED (From - To)</b> 1September2011-31August2012	
<b>4. TITLE AND SUBTITLE</b> A Prosthesis to Train the Proprioceptive Capabilities of the Residual Limb of Military Personnel Recovering From Lower Limb Amputation				<b>5a. CONTRACT NUMBER</b>	
				<b>5b. GRANT NUMBER</b> W81XWH-10-1-0573	
				<b>5c. PROGRAM ELEMENT NUMBER</b>	
<b>6. AUTHOR(S)</b>  Brian Glaister				<b>5d. PROJECT NUMBER</b>	
				<b>5e. TASK NUMBER</b>	
				<b>5f. WORK UNIT NUMBER</b>	
<b>7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES)</b> Cadence Biomedical Seattle, WA 98103				<b>8. PERFORMING ORGANIZATION REPORT NUMBER</b>	
<b>9. SPONSORING / MONITORING AGENCY NAME(S) AND ADDRESS(ES)</b> USAMRMC Fort Detrick, MD 21702-5012				<b>10. SPONSOR/MONITOR'S ACRONYM(S)</b>	
				<b>11. SPONSOR/MONITOR'S REPORT NUMBER(S)</b>	
<b>12. DISTRIBUTION / AVAILABILITY STATEMENT</b> Approved for public release, distribution unlimited					
<b>13. SUPPLEMENTARY NOTES</b>					
<b>14. ABSTRACT</b> Learning to walk following lower limb amputation takes many months owing largely to the fact that new amputees cannot perceive when the prosthetic foot is in contact with the ground. To overcome this limitation, we have developed a system that detects foot pressures and relays this information to the residual limb. To data, a prototype sensor/stimulator system was been developed and we are preparing for human subjects tests.					
<b>15. SUBJECT TERMS</b> Exotendon, orthosis, biomechanics, gait					
<b>16. SECURITY CLASSIFICATION OF:</b> U			<b>17. LIMITATION OF ABSTRACT</b>  UU	<b>18. NUMBER OF PAGES</b>  4	<b>19a. NAME OF RESPONSIBLE PERSON</b> USAMRMC
<b>a. REPORT</b> U	<b>b. ABSTRACT</b> U	<b>c. THIS PAGE</b> U			<b>19b. TELEPHONE NUMBER (include area code)</b>

## Introduction

One of the major obstacles amputees must overcome during post-operative rehabilitation is learning to maintain balance. Over time, amputees learn to substitute tactile feedback discerned from pressures imparted on the residual limb by the prosthetic socket to control balance [1], but during the post-operative period the lack of afferent feedback from the ankle and foot poses challenge. To compensate for the lack of afferent feedback, amputees look down at the prosthesis to sense body position [1, 2] and continue to do so for up to eight months post-surgery [1]. Visual dependence is problematic since it may cause amputees to miss obstacles that need to be avoided. Furthermore, the visual feedback is not as effective as somatosensory feedback. As such, falls are frequent [3] which can cause injury to the surgical site, further delaying the rehabilitation process [4] and increasing costs.

In an effort to improve post-operative gait training for lower limb amputees, we have developed a vibrotactile sensory feedback system (Fig 1). This system uses sensors placed on the prosthetic foot to measure loading, process this information, and output it to vibrotactile stimulators that communicate both the level and location of the load. It is our hope that by including this capability in a prosthesis, we may be able to help new amputees learn to walk in a more timely manner.




Fig 1: Supplementary Feedback System

## Body

The research proposed for this project included three specific aims: 1) Design and fabricate a vibrotactile sensory feedback system, 2) Investigate appropriate body sites for sensory feedback, and 3) Investigate the ability of the system to help amputees modulate body position in response to sensory feedback.

To date, we have developed a multi-channel sensory feedback device. The device uses an external power supply and is controlled by an Atmel microprocessor. Inputs are taken from Tekscan Flexiforce sensors which output a voltage proportional to the pressure present in the sensors. The sensor output is processed by the microprocessor, and then the appropriate voltage is released to the vibrating stimulator. Vibration amplitude is proportional to sensor pressure.

We had secured IRB approval to begin recruiting for this study, but we moved our office and the IRB has instructed us to submit a modification to our protocol with our new address. This modification is now pending and once approved we will recruit the lower limb amputee subjects and complete the study.

## Key Research Accomplishments

- Completion of multi-input sensory feedback device.

## Reportable Outcomes

A pending provisional patent application reported in the last report was tabled for budgetary reasons. The application will be written and filed prior to public disclosure of patentable mechanisms in the design.

## Conclusion

We have developed a new prototype sensory feedback device. Data collection is pending.

## References

1. Geurts, A.C., et al., *Postural reorganization following lower limb amputation. Possible motor and sensory determinants of recovery*. Scand J Rehabil Med, 1992. **24**(2): p. 83-90.
2. Gauthier-Gagnon, C., et al., *Augmental sensory feedback in the early training of standing balance of below-knee amputees*. Physiotherapy Canada, 1986. **38**: p. 137-42.

3. Quai, T.M., S.G. Brauer, and J.C. Nitz, *Somatosensation, circulation and stance balance in elderly dysvascular transtibial amputees*. Clin Rehabil, 2005. **19**(6): p. 668-76.
4. Goldberg, T., S. Goldberg, and J. Pollak, *Postoperative management of lower extremity amputation*. Phys Med Rehabil Clin N Am, 2000. **11**(3): p. 559-68, vi.